



Proficiency Testing of Conducted Emission Measurements PTC(CE-9k-30M-VIII)

Firenze, January 15, 2025

Rev. 0

C. Carobbi and A. Bonci, University of Florence, Firenze, Italy

Travelling Sample for the 9 kHz to 30 MHz frequency range (Conducted Emission)



General information

- Number of participants: 19
- Start date: November 2023
- Stop date: December 2024
- Scheme of the proficiency test PTC(CE-9k-30M-VII):
<https://www.dinfo.unifi.it/vp-436-schemes-of-the-proficiency-tests.html>
- Issues faced:
 - Some laboratories issued the report (annex C) with a severe delay (several months) with respect the expected date (1 week after measurements).
 - 1 multi-site laboratory was severely delayed in performing measurements with respect to the scheduled weeks (nearly 3 months). The PT closed with the same delay.

Measurement procedure

- Voltage measurement by using the AMN and EMI receiver is preceded by a preliminary check of one harmonic generated by the Sample.
- Measurement by using the AMN and EMI receiver are performed according to §7.4.2 of EN 55016-2-1:2014 and next amendments, by using a V-type Artificial Mains Network (AMN).
- The Laboratory measures the amplitude of ten (10) harmonics selected by the Coordinator in the frequency range between 9 kHz and 30 MHz (i.e., covering both band A and band B). The disturbance injected by the Sample in line and neutral conductors is measured. A total number of twenty (20) measurements (two conductors times ten frequencies) is reported to the Coordinator by the Laboratory.

Sequence of operations

- Connect the Sample to the EUT port of the AMN;
- Power up the AMN;
- Measure the amplitude of the ten harmonics selected by the Coordinator by using the EMI receiver set with average detector;
- Power off the AMN;
- Disconnect the Sample from the AMN.

Measurement result

- The measurement result provided by the Laboratory consists of:
 - The estimate x , expressed in dB(μ V), of the amplitude of the selected harmonics, measured both line-to-ground (x_{line}) and neutral-to-ground ($x_{neutral}$);
 - The expanded uncertainty of the estimate x , U_{lab} , expressed in dB and obtained multiplying the combined standard uncertainty by the coverage factor $k = 2$ (which corresponds to a coverage probability of about 95 % assuming normal distribution).

Reference values

Reference
values

Statistical analysis of
measurement results

\bar{x}^* , s^*

Statistical (robust) analysis

$x_1, x_2, \dots, x_i, \dots, x_p$ } Raw data (p participants)

$x^* = \text{median of } x_i \quad (i = 1, 2, \dots, p)$

$s^* = 1,483 \text{ median of } |x_i - x^*| \quad (i = 1, 2, \dots, p)$

} Initial reference value

$\delta = 1,5s^*$

$x_i^* = \begin{cases} x^* - \delta, & \text{if } x_i < x^* - \delta \\ x^* + \delta, & \text{if } x_i > x^* + \delta \\ x_i, & \text{otherwise} \end{cases}$

} Transformed set of data

$x^* = \sum x_i^* / p$

$s^* = 1,134 \sqrt{\sum (x_i^* - x^*)^2 / (p - 1)}$

} New reference value
(iterative algorithm)



Performance statistic ζ (Participant)

- Performance statistic ζ (clause 9.6 of ISO 13528:2015) that the Coordinator applies to the Participant providing the measurement result x_i with standard uncertainty u_{x_i}


$$\zeta_i = \frac{x_i - X}{\sqrt{u_{x_i}^2 + u_X^2}} \quad X = x^*, u_X = \frac{1,25 \cdot s^*}{\sqrt{p}}$$


$$\begin{cases} 2 < |\zeta_i| < 3 \Rightarrow \text{warning} \\ 3 < |\zeta_i| \Rightarrow \text{action} \end{cases}$$


Results


Harmonic code to frequency conversion


Band	Harmonic #	Frequency MHz
A	2	0.0195
A	10	0.0819
A	17	0.1365
B	1	0.5000
B	8	7.500
B	12	11.500
B	17	16.500
B	21	20.500
B	25	24.500
B	29	28.500























Reproducibility as quantified by robust standard deviation s^*

f MHz	s^* dB
0.0195	1.1
0.0819	0.5
0.1365	0.3
0.5	0.4
7.5	0.4
11.5	0.5
16.5	0.4
20.5	0.7
24.5	0.7
28.5	0.7

Remarks

- The measurement results provided by the 19 participants at the 10 measurement frequencies selected by the Coordinator are approximately within -2 dB to $+14$ dB from the reference values. Most of measurement results are within -2 dB to $+2$ dB from the reference values.
- 374 measurement results were provided by the participants and 6 signals (2 warning and 4 action) were issued.
- The standard measurement uncertainty declared by the laboratories is comprised approximately between 0.6 dB and 1.9 dB, robust standard deviation s^* is less than 1.1 dB.